

U.S. FDA Transition Protocol Test Matrix— Nelson Laboratories Testing Status (as of May 13, 2015)



		Style	Pouches and Bags						Form-Fill-Seal						Rigid Trays								
EO	Coated	1073B	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21
EO	Uncoated	1073B	22	23	24	25	26	27															
Gamma	Coated	1073B	28		29		30	31	32		33	34	35	36	37		38	39					
Gamma	Uncoated	1073B	40		41		42																
Electron-beam	Coated	1073B							43	44	45												
Electron-beam	Uncoated	1073B	46		47		48																
EO	Coated	1059B							49	50	51												
EO	Uncoated	1059B	52	53	54	55	56	57	58	59	60												

LEGEND (% COMPLETE):

- 100% PRE-, POST-, 1, 3 & 5 YEAR ACC. AGING AND 1 YEAR REAL-TIME AGING DATA COMPLETE
- 50% PRE-, POST-, 1, 3, 5 & 7 YEAR ACC. AGING AND 1 YEAR REAL-TIME AGING DATA COMPLETE
- In Progress PRE-, POST-, 1, 3, 5, 7 & 10 YEAR ACC. AGING AND 1 YEAR REAL-TIME AGING DATA COMPLETE
- In Progress PRE-, POST-, 1, 3 & 5 YEAR ACC. AGING AND 1 & 3 YEAR REAL-TIME AGING DATA COMPLETE